

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
 Data File : VL041676.D
 Acq On : 7 Nov 2024 11:34
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Quant Time: Nov 08 00:10:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Nov 08 00:07:44 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 11/09/2024
 Supervised By :Semsettin Yesilyurt 11/09/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	1184622	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	3332226	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	3067671	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.773	95	2285234	10.073	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.700%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.726	85	181618	0.848	ppbv	98
3) Chlorodifluoromethane	2.671	51	164131	1.063	ppbv	100
4) Chloromethane	2.812	50	86683	1.060	ppbv	100
5) Vinyl Chloride	2.915	62	78147	1.055	ppbv	100
6) Bromomethane	3.118	94	38811	0.995	ppbv	89
7) Chloroethane	3.195	64	27451	1.023	ppbv	94
8) Dichlorotetrafluoroethane	2.854	85	198669	1.039	ppbv	99
9) Propene	2.690	41	72231	1.074	ppbv	95
10) Heptane	7.574	43	148906	0.988	ppbv	96
11) Trichlorofluoromethane	3.562	101	194775	1.044	ppbv	98
12) 1,1,2-Trichlorotrifluo...	4.063	101	141527	1.044	ppbv	98
13) Ethanol	3.256	45	3961	0.947	ppbv #	48
14) Bromoethene	3.359	108	55094	1.011	ppbv #	92
15) Acetone	3.475	43	166255	1.140	ppbv	99
16) 1,3-Butadiene	2.980	39	67777m	0.927	ppbv	
17) tert-Butyl alcohol	3.883	59	153265	1.198	ppbv	99
18) 1,1-Dichloroethene	3.873	96	58180	0.979	ppbv	97
19) Isopropyl Alcohol	3.584	45	82398	1.168	ppbv #	91
20) Methylene Chloride	3.928	84	59394	1.107	ppbv	96
21) Allyl Chloride	3.992	41	95568	1.061	ppbv	98
22) trans-1,2-Dichloroethene	4.446	96	60981m	1.028	ppbv	
23) Vinyl Acetate	4.626	43	66087m	0.964	ppbv	
24) 1,1-Dichloroethane	4.555	63	132214	1.002	ppbv	99
25) Ethyl Acetate	5.205	43	316814	1.050	ppbv	98
26) Hexane	5.208	57	122113	1.053	ppbv	94
27) Carbon Disulfide	4.121	76	125113	0.961	ppbv	97
28) Methyl tert-Butyl Ether	4.587	73	79309	1.062	ppbv	98
29) Chloroform	5.269	83	197891	1.034	ppbv	100
30) Cyclohexane	6.597	84	88127	0.946	ppbv	89
31) cis-1,2-Dichloroethene	5.073	61	115115	1.032	ppbv	94
32) 1,1,1-Trichloroethane	5.996	97	186425	1.063	ppbv	100
34) 2-Butanone	4.790	43	166525	1.004	ppbv	99
35) Carbon Tetrachloride	6.491	117	187475	1.064	ppbv	96
36) Benzene	6.369	78	250318	1.014	ppbv	98
37) 1,2-Dichloroethane	5.803	62	144449	1.037	ppbv	100
38) Trichloroethene	7.288	130	96191	1.010	ppbv	93
39) 1,2-Dichloropropane	7.069	63	99780	1.023	ppbv	91
40) 1,4-Dioxane	7.311	88	35039m	1.014	ppbv	
41) Tetrahydrofuran	5.571	42	93643	0.989	ppbv	99
42) Bromodichloromethane	7.246	83	188568	0.974	ppbv	100
43) Methyl Methacrylate	7.478	69	78902	0.910	ppbv	90
44) 2,2,4-Trimethylpentane	7.327	57	435343	1.041	ppbv	98
45) t-1,3-Dichloropropene	8.709	75	70088m	0.904	ppbv	

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
 Data File : VL041676.D
 Acq On : 7 Nov 2024 11:34
 Operator : SY/MD
 Sample : VSTDICC001
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 11/09/2024
 Supervised By :Semsettin Yesilyurt 11/09/2024

Quant Time: Nov 08 00:10:33 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri Nov 08 00:07:44 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	8.143	75	101425	0.901	ppbv	95
47) 1,1,2-Trichloroethane	8.896	97	104883	0.997	ppbv	92
48) Dibromochloromethane	9.709	129	152757	0.952	ppbv	99
49) Bromoform	12.400	173	127488m	0.931	ppbv	
50) 4-Methyl-2-Pentanone	8.188	43	250885	0.985	ppbv	99
51) 2-Hexanone	9.571	43	165158	0.892	ppbv #	100
52) Tetrachloroethene	10.610	164	92977	1.102	ppbv	93
53) Toluene	9.217	91	253501	0.976	ppbv	98
54) 1,2-Dibromoethane	10.008	107	122096	0.994	ppbv	97
56) 1,1,1,2-Tetrachloroethane	11.510	131	133283	1.043	ppbv #	1
57) Chlorobenzene	11.523	112	239417	1.047	ppbv	97
58) Ethyl Benzene	12.089	91	334939	0.986	ppbv	91
59) m/p-Xylene	12.368	91	625077m	2.056	ppbv	
60) o-Xylene	13.053	91	293793	1.012	ppbv	100
61) Styrene	12.892	104	102947	0.859	ppbv	96
62) Isopropylbenzene	14.031	105	468382	1.045	ppbv	98
63) 1,1,2,2-Tetrachloroethane	13.047	83	226867	1.074	ppbv	97
64) n-propylbenzene	14.915	120	115677	0.987	ppbv	82
65) tert-Butylbenzene	16.092	119	424019	1.053	ppbv	98
66) Benzyl Chloride	16.326	91	26944m	0.929	ppbv	
67) sec-Butylbenzene	16.638	105	596449	1.051	ppbv	99
69) p-Isopropyltoluene	16.998	119	480893	1.048	ppbv	99
70) n-Butylbenzene	17.889	91	484367	1.030	ppbv	97
71) 2-Chlorotoluene	14.799	91	331921	1.024	ppbv	99
72) 4-Ethyltoluene	15.191	105	327541	0.972	ppbv	98
73) 1,3,5-Trimethylbenzene	15.352	105	290287	0.982	ppbv	93
74) 1,2,4-Trimethylbenzene	16.111	105	360083	1.047	ppbv #	89
75) 1,3-Dichlorobenzene	16.326	146	218228	1.028	ppbv	99
76) 1,4-Dichlorobenzene	16.471	146	203100	0.973	ppbv	95
77) 1,2-Dichlorobenzene	17.143	146	221404	1.050	ppbv	98
78) Hexachloro-1,3-Butadiene	22.632	225	190183m	1.140	ppbv	
79) Naphthalene	21.394	128	226301m	1.070	ppbv	
80) Naphthalene,2-methyl-	24.374	142	39140m	0.998	ppbv	
81) 1,2,4-Trichlorobenzene	21.146	180	138641m	1.054	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
Data File : VL041676.D
Acq On : 7 Nov 2024 11:34
Operator : SY/MD
Sample : VSTDICC001
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 11/09/2024
Supervised By :Semsettin Yesilyurt 11/09/2024

Quant Time: Nov 08 00:10:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug
QLast Update : Fri Nov 08 00:07:44 2024
Response via : Initial Calibration

