

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
 Data File : VL041690.D
 Acq On : 7 Nov 2024 23:03
 Operator : SY/MD
 Sample : P4696-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA1DL

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 11/09/2024
 Supervised By :Semsettin Yesilyurt 11/09/2024

Quant Time: Nov 08 03:54:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Fri Nov 08 01:01:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.185	49	1071435	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	2948814	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	2677366	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.770	95	1941685	9.843	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.400%
Target Compounds						
					Qvalue	
13) Ethanol	3.234	45	212975m	57.180	ppbv	
15) Acetone	3.468	43	203008m	1.540	ppbv	
19) Isopropyl Alcohol	3.568	45	2508130	39.795	ppbv #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
Data File : VL041690.D
Acq On : 7 Nov 2024 23:03
Operator : SY/MD
Sample : P4696-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA1DL

Quant Time: Nov 08 03:54:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Nov 08 01:01:00 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 11/09/2024
Supervised By : Semsettin Yesilyurt 11/09/2024

