

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
 Data File : VL041694.D
 Acq On : 8 Nov 2024 1:39
 Operator : SY/MD
 Sample : CAN 10154
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10154

Quant Time: Nov 08 03:55:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Nov 08 01:01:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	991197	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	2720698	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.462	117	2387368	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.770	95	1790227	10.177	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.800%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110724\
Data File : VL041694.D
Acq On : 8 Nov 2024 1:39
Operator : SY/MD
Sample : CAN 10154
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10154


Quant Time: Nov 08 03:55:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Nov 08 01:01:00 2024
Response via : Initial Calibration








#1
 Bromochloromethane
 Concen: 10.000 ppbv
 RT: 5.192 min Scan# 804
 Delta R.T. 0.004 min
 Lab File: VL041694.D
 Acq: 8 Nov 2024 1:39

Instrument :
 MSVOA_L
 ClientSampleId :
 CAN 10154







Tgt Ion: 49 Resp: 991197
 Ion Ratio Lower Upper
 49 100
 128 47.4 23.8 71.2
 130 59.8 30.1 90.5



#33
 1,4-Difluorobenzene
 Concen: 10.000 ppbv
 RT: 6.648 min Scan# 1258
 Delta R.T. -0.003 min
 Lab File: VL041694.D
 Acq: 8 Nov 2024 1:39

Tgt Ion: 114 Resp: 2720698
 Ion Ratio Lower Upper
 114 100
 63 25.1 20.3 30.5
 88 17.5 14.2 21.4





#55

Chlorobenzene-d5

Concen: 10.000 ppbv

RT: 11.462 min Scan# 27

Delta R.T. -0.003 min

Lab File: VL041694.D

Acq: 8 Nov 2024 1:39

Instrument :

MSVOA_L

ClientSampleId :

CAN 10154





Tgt Ion:117 Resp: 2387368

Ion Ratio Lower Upper

117 100

82 64.9 49.8 74.8

119 33.3 24.2 36.4



#68

1-Bromo-4-Fluorobenzene

Concen: 10.177 ppbv

RT: 13.770 min Scan# 3473

Delta R.T. -0.003 min

Lab File: VL041694.D

Acq: 8 Nov 2024 1:39

Tgt Ion: 95 Resp: 1790227

Ion Ratio Lower Upper

95 100

174 70.2 55.6 83.4

175 4.9 4.2 6.2

