

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110823\
 Data File : VL040810.D
 Acq On : 8 Nov 2023 14:10
 Operator : SY/MD
 Sample : VL1108ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1108ABL01

Quant Time: Nov 09 00:29:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101123AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Nov 09 00:25:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	1250726	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.658	114	3208466	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.458	117	2766898	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.770	95	2058480	10.111	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110823\
Data File : VL040810.D
Acq On : 8 Nov 2023 14:10
Operator : SY/MD
Sample : VL1108ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1108ABL01

Quant Time: Nov 09 00:29:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101123AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Nov 09 00:25:02 2023
Response via : Initial Calibration

