

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110922\
 Data File : VL039845.D
 Acq On : 9 Nov 2022 12:36
 Operator : SY/AP
 Sample : VL1109ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1109ABL01

Quant Time: Nov 10 03:47:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Thu Nov 03 09:27:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.173	49	925919	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.623	114	2510981	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.430	117	2235817	10.000	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.735	95	1771324	10.585	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110922\
Data File : VL039845.D
Acq On : 9 Nov 2022 12:36
Operator : SY/AP
Sample : VL1109ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1109ABL01

Quant Time: Nov 10 03:47:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Thu Nov 03 09:27:12 2022
Response via : Initial Calibration

