

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110922\
 Data File : VL039855.D
 Acq On : 9 Nov 2022 19:31
 Operator : SY/AP
 Sample : N5527-03DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 10604-OUTSIDEDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/10/2022
 Supervised By :Mahesh Dadoda 11/11/2022

Quant Time: Nov 10 03:51:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Nov 03 09:27:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.175	49	999805	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	6.629	114	2315444	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.432	117	2096588	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.741	95	1680569	10.709	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	= 107.100%	
Target Compounds						
						Qvalue
3) Chlorodifluoromethane	2.680	51	18660	0.157	ppbv	100
13) Ethanol	3.253	45	22695m	5.171	ppbv	
15) Acetone	3.475	43	44143	0.363	ppbv	90
20) Methylene Chloride	3.931	84	89856	2.064	ppbv	90
26) Hexane	5.201	57	88895	0.795	ppbv #	40
59) m/p-Xylene	12.307	91	102337m	0.432	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL110922\
Data File : VL039855.D
Acq On : 9 Nov 2022 19:31
Operator : SY/AP
Sample : N5527-03DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
10604-OUTSIDEDL

Quant Time: Nov 10 03:51:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Nov 03 09:27:12 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/10/2022
Supervised By :Mahesh Dadoda 11/11/2022

