

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111020\
 Data File : VL035904.D
 Acq On : 11 Nov 2020 8:31
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 11/12/2020 9:03:22 AM

Quant Time: Nov 11 16:53:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 0
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1112721	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3972356	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3632004	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2735924m	10.60	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.86	43	34018m	0.286	ppbv	
20) Methylene Chloride	4.34	84	14696	0.175	ppbv	95
35) Carbon Tetrachloride	7.01	117	8492	0.040	ppbv #	94
52) Tetrachloroethene	11.21	164	5228m	0.037	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111020\
Data File : VL035904.D
Acq On : 11 Nov 2020 8:31
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
11/12/2020 9:03:22 AM

Quant Time: Nov 11 16:53:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

