


Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111020\
 Data File : VL035891.D
 Acq On : 10 Nov 2020 15:10
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110320AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
 Last Update : Wed Nov 04 02:40:06 2020



AutoFind: Scans 3614, 3615, 3616; Background Corrected with Scan 3591

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	16.6	183082	PASS
75	95	30	66	48.5	534506	PASS
95	95	100	100	100.0	1101994	PASS
96	95	5	9	6.8	75298	PASS
173	174	0.00	2	0.6	4577	PASS
174	95	50	120	73.5	809600	PASS
175	174	4	9	7.3	58954	PASS
176	174	93	101	96.9	784192	PASS
177	176	5	9	6.8	53333	PASS