

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111021\
 Data File : VL038019.D
 Acq On : 11 Nov 2021 8:54
 Operator : SY/AP
 Sample : OC111021 CAN 10060
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 QC111021 CAN 10060

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/12/2021
 Supervised By :Mahesh Dadoda 11/17/2021

Quant Time: Nov 12 05:59:18 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110121A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Nov 02 05:54:07 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1228164	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.15	114	3774740	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	3256962	10.00	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.38	95	2301674	9.72	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.20%
Target Compounds						
38) Trichloroethene	7.81	130	23784m	0.186	ppbv	Ovalue
52) Tetrachloroethene	11.19	164	110634	0.949	ppbv	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111021\
Data File : VL038019.D
Acq On : 11 Nov 2021 8:54
Operator : SY/AP
Sample : OC111021 CAN 10060
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC111021 CAN 10060

Quant Time: Nov 12 05:59:18 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110121
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/12/2021
Supervised By :Mahesh Dadoda 11/17/2021

