

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111021\
 Data File : VL038020.D
 Acq On : 11 Nov 2021 9:31
 Operator : SY/AP
 Sample : M4598-04DL2 1200X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 356-4DL2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/12/2021
 Supervised By :Mahesh Dadoda 11/17/2021

Quant Time: Nov 15 07:44:16 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110121A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Nov 02 05:54:07 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1264791	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.15	114	3939242	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	3467268	10.00	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.38	95	2446672	9.71	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.10%
Target Compounds						
31) cis-1,2-Dichloroethene	5.52	61	23979m	0.179	ppbv	
38) Trichloroethene	7.81	130	679719	5.105	ppbv	97
52) Tetrachloroethene	11.20	164	3455766	28.397	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111021\
Data File : VL038020.D
Acq On : 11 Nov 2021 9:31
Operator : SY/AP
Sample : M4598-04DL2 1200X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
356-4DL2

Manual Integrations
APPROVED

Quant Time: Nov 15 07:44:16 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110121
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Nov 02 05:54:07 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 11/12/2021
Supervised By :Mahesh Dadoda 11/17/2021

