


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL111021\
 Data File : VL038006.D
 Acq On : 10 Nov 2021 10:33
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL032521AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Thu Mar 25 09:42:40 2021
 Last Update : Thu Mar 25 09:42:40 2021



AutoFind: Scans 3611, 3612, 3613; Background Corrected with Scan 3579

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	21.8	195392	PASS
75	95	30	66	50.7	454229	PASS
95	95	100	100	100.0	896512	PASS
96	95	5	9	7.0	62752	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	70.0	627712	PASS
175	174	4	9	8.9	55720	PASS
176	174	93	101	96.6	606122	PASS
177	176	5	9	6.6	39749	PASS