

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111022\
 Data File : VL039870.D
 Acq On : 10 Nov 2022 19:14
 Operator : SY/AP
 Sample : N5579-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-ET1DL

Quant Time: Nov 11 05:02:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Nov 03 09:27:12 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 11/11/2022
 Supervised By : Mahesh Dadoda 11/11/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	992657	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.648	114	2268812	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.455	117	1980553	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.764	95	1625046	10.962	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	= 109.600%	
Target Compounds						
						Qvalue
13) Ethanol	3.269	45	26273m	6.029	ppbv	
52) Tetrachloroethene	10.606	164	118736	1.656	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111022\
Data File : VL039870.D
Acq On : 10 Nov 2022 19:14
Operator : SY/AP
Sample : N5579-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-ET1DL

Quant Time: Nov 11 05:02:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Nov 03 09:27:12 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/11/2022
Supervised By : Mahesh Dadoda 11/11/2022

