

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111022\
 Data File : VL039882.D
 Acq On : 11 Nov 2022 9:00
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 11/11/2022
 Supervised By :Mahesh Dadoda 11/11/2022

Quant Time: Nov 11 09:32:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Thu Nov 03 09:27:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.201	49	993228	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.658	114	2291613	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.465	117	1984353	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.773	95	1588949	10.698	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.000%
Target Compounds						
					Qvalue	
20) Methylene Chloride	3.947	84	4817	0.111	ppbv #	82
35) Carbon Tetrachloride	6.500	117	4981	0.037	ppbv #	82
38) Trichloroethene	7.291	130	7247m	0.070	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111022\
Data File : VL039882.D
Acq On : 11 Nov 2022 9:00
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 11/11/2022
Supervised By :Mahesh Dadoda 11/11/2022

Quant Time: Nov 11 09:32:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Thu Nov 03 09:27:12 2022
Response via : Initial Calibration

