

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111325\
 Data File : VL043192.D
 Acq On : 13 Nov 2025 09:47
 Operator : SY/MD
 Sample : VL1113ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1113ABL01

Quant Time: Nov 14 04:36:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Wed Oct 15 02:12:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.794	49	136642	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.968	114	369240	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.888	117	306160	10.000	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.383	95	228486	9.601	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111325\
Data File : VL043192.D
Acq On : 13 Nov 2025 09:47
Operator : SY/MD
Sample : VL1113ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1113ABL01

Quant Time: Nov 14 04:36:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL101425AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Wed Oct 15 02:12:58 2025
Response via : Initial Calibration

