

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111423\
 Data File : VL040862.D
 Acq On : 14 Nov 2023 22:29
 Operator : SY/MD
 Sample : 05275-06DL 2X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NT-IA-6DL

Quant Time: Nov 16 04:53:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Nov 13 15:42:46 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 11/17/2023
 Supervised By :Semsettin Yesilyurt 11/17/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	986516	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.639	114	2611987	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.436	117	2419715	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.744	95	1786996	9.433	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.300%
Target Compounds						
					Qvalue	
15) Acetone	3.484	43	460617	4.384	ppbv	91
20) Methylene Chloride	3.941	84	146964	3.389	ppbv	97
26) Hexane	5.211	57	851366	8.842	ppbv #	40
36) Benzene	6.356	78	61273	0.289	ppbv	97
38) Trichloroethene	7.275	130	9946m	0.121	ppbv	
53) Toluene	9.195	91	134138	0.575	ppbv	96
58) Ethyl Benzene	12.063	91	30162m	0.096	ppbv	
59) m/p-Xylene	12.314	91	73377m	0.273	ppbv	
60) o-Xylene	13.027	91	27771m	0.109	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111423\
Data File : VL040862.D
Acq On : 14 Nov 2023 22:29
Operator : SY/MD
Sample : 05275-06DL 2X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NT-IA-6DL

Quant Time: Nov 16 04:53:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Nov 13 15:42:46 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 11/17/2023
Supervised By : Semsettin Yesilyurt 11/17/2023

