

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
 Data File : VL040868.D
 Acq On : 15 Nov 2023 14:04
 Operator : SY/MD
 Sample : 05277-01 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NT-SS-2

Quant Time: Nov 16 05:52:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Nov 13 15:42:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	1083008	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.642	114	2783950	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.442	117	2481101	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.748	95	1919579	9.882	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.800%
Target Compounds						
					Qvalue	
15) Acetone	3.481	43	1611069	13.967	ppbv	93
20) Methylene Chloride	3.944	84	62143	1.305	ppbv	95
26) Hexane	5.211	57	167287	1.583	ppbv #	43
27) Carbon Disulfide	4.134	76	305616	2.280	ppbv	98
36) Benzene	6.362	78	140966	0.624	ppbv	96
53) Toluene	9.201	91	78299	0.315	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
Data File : VL040868.D
Acq On : 15 Nov 2023 14:04
Operator : SY/MD
Sample : 05277-01 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NT-SS-2

Quant Time: Nov 16 05:52:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Nov 13 15:42:46 2023
Response via : Initial Calibration

