

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
 Data File : VL040869.D
 Acq On : 15 Nov 2023 15:08
 Operator : SY/MD
 Sample : 05277-01DUP 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NT-SS-2DUP

Quant Time: Nov 21 03:22:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Nov 20 03:50:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	1058431	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.642	114	2717865	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.439	117	2439493	10.000	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.747	95	1926560	10.087	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.900%

Target Compounds						Qvalue
15) Acetone	3.481	43	1562388	13.860	ppbv	92
20) Methylene Chloride	3.944	84	136229	2.928	ppbv	95
26) Hexane	5.211	57	204247	1.977	ppbv #	40
27) Carbon Disulfide	4.134	76	289307	2.209	ppbv	98
36) Benzene	6.359	78	139473	0.632	ppbv	94
53) Toluene	9.198	91	73741	0.304	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
Data File : VL040869.D
Acq On : 15 Nov 2023 15:08
Operator : SY/MD
Sample : 05277-01DUP 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NT-SS-2DUP

Quant Time: Nov 21 03:22:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Nov 20 03:50:24 2023
Response via : Initial Calibration

