

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
 Data File : VL040874.D
 Acq On : 15 Nov 2023 18:52
 Operator : SY/MD
 Sample : 05277-06 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NT-SS-8

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/20/2023
 Supervised By :Mahesh Dadoda 11/21/2023

Quant Time: Nov 16 05:55:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2
 QLast Update : Mon Nov 13 15:42:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.195	49	1028212	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.648	114	2534361	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.446	117	2273677	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.754	95	1746036	9.808	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.100%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.761	85	22542	0.108	ppbv	96
13) Ethanol	3.269	45	274009	25.506	ppbv #	100
15) Acetone	3.491	43	754322	6.888	ppbv	100
20) Methylene Chloride	3.951	84	777544m	17.205	ppbv	
26) Hexane	5.218	57	752381	7.497	ppbv #	41
53) Toluene	9.205	91	47924	0.212	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
Data File : VL040874.D
Acq On : 15 Nov 2023 18:52
Operator : SY/MD
Sample : 05277-06 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NT-SS-8

Quant Time: Nov 16 05:55:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Nov 13 15:42:46 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/20/2023
Supervised By :Mahesh Dadoda 11/21/2023

