

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
 Data File : VL040885.D
 Acq On : 16 Nov 2023 1:54
 Operator : SY/MD
 Sample : 05277-08 2X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 NT-SS-9

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/20/2023
 Supervised By :Mahesh Dadoda 11/21/2023

Quant Time: Nov 16 06:36:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Mon Nov 13 15:42:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.205	49	920925	10.000	ppbv	0.03
33) 1,4-Difluorobenzene	6.655	114	2264902	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.449	117	2007624	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.757	95	1588560	10.106	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.100%
Target Compounds						
					Qvalue	
13) Ethanol	3.275	45	166351	17.289	ppbv #	100
15) Acetone	3.504	43	447788	4.565	ppbv	98
20) Methylene Chloride	3.960	84	49541	1.224	ppbv	93
26) Hexane	5.227	57	170701	1.899	ppbv #	40
36) Benzene	6.378	78	30663	0.167	ppbv	97
38) Trichloroethene	7.295	130	4923m	0.069	ppbv	
53) Toluene	9.208	91	54790	0.271	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL111523\
Data File : VL040885.D
Acq On : 16 Nov 2023 1:54
Operator : SY/MD
Sample : 05277-08 2X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
NT-SS-9

Quant Time: Nov 16 06:36:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111323AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Mon Nov 13 15:42:46 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 11/20/2023
Supervised By : Mahesh Dadoda 11/21/2023

