

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111620\
 Data File : VL035944.D
 Acq On : 16 Nov 2020 15:18
 Operator : SY/AP
 Sample : L4755-01 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AS-EFF-201112

Quant Time: Nov 17 09:59:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1282738	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4426514	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	4192461	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	3091734	10.38	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	2.99	51	26023	0.169	ppbv	94
11) Trichlorofluoromethane	3.95	101	127773	0.421	ppbv	94
13) Ethanol	3.63	45	8943	0.733	ppbv	73
15) Acetone	3.87	43	33049	0.241	ppbv #	41
20) Methylene Chloride	4.35	84	46540	0.482	ppbv	90
26) Hexane	5.69	57	15897	0.107	ppbv #	45
31) cis-1,2-Dichloroethene	5.55	61	45307	0.296	ppbv	90
38) Trichloroethene	7.84	130	432249	2.689	ppbv	99
52) Tetrachloroethene	11.23	164	284779	1.827	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111620\
Data File : VL035944.D
Acq On : 16 Nov 2020 15:18
Operator : SY/AP
Sample : L4755-01 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AS-EFF-201112

Quant Time: Nov 17 09:59:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

