

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111620\
 Data File : VL035946.D
 Acq On : 16 Nov 2020 16:37
 Operator : SY/AP
 Sample : L4721-06DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LA-6DL

Quant Time: Nov 17 10:39:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1293317	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4472501	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	4218911	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	3059473	10.21	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.10%

Target Compounds					Ovalue
9) Propene	3.02	41	11700	0.206 ppbv	87
13) Ethanol	3.62	45	28356	2.306 ppbv	94
15) Acetone	3.86	43	87152	0.630 ppbv #	79
19) Isopropyl Alcohol	3.99	45	13974m	0.164 ppbv	
20) Methylene Chloride	4.35	84	97763	1.004 ppbv	95
26) Hexane	5.69	57	81899	0.545 ppbv #	44
53) Toluene	9.81	91	95914	0.239 ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111620\
Data File : VL035946.D
Acq On : 16 Nov 2020 16:37
Operator : SY/AP
Sample : L4721-06DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LA-6DL

Quant Time: Nov 17 10:39:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

