

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111620\
 Data File : VL035948.D
 Acq On : 16 Nov 2020 17:56
 Operator : SY/AP
 Sample : L4722-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SR-1DL

Quant Time: Nov 17 21:52:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1275402	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4388597	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	4138270	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	3092403	10.52	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	2179538	179.716	ppbv	95
15) Acetone	3.85	43	365609	2.681	ppbv	99
19) Isopropyl Alcohol	3.98	45	383372	4.575	ppbv #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111620\
Data File : VL035948.D
Acq On : 16 Nov 2020 17:56
Operator : SY/AP
Sample : L4722-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SR-1DL

Quant Time: Nov 17 21:52:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

