

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111620\
 Data File : VL035954.D
 Acq On : 16 Nov 2020 21:52
 Operator : SY/AP
 Sample : L4722-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SR-4DL

Manual Integrations
 APPROVED

MMDadoda
 11/18/2020 9:07:52 AM

Quant Time: Nov 17 22:14:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 0
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1283073	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4381931	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	4129289	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.42	95	3066368	10.45	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	2.98	51	115375	0.751	ppbv #	88
9) Propene	3.02	41	20211	0.359	ppbv	85
10) Heptane	8.12	43	27668	0.178	ppbv	94
13) Ethanol	3.60	45	2385051	195.486	ppbv	95
15) Acetone	3.84	43	1978348	14.422	ppbv	98
19) Isopropyl Alcohol	3.97	45	459475	5.451	ppbv #	90
59) m/p-Xylene	12.97	91	69729m	0.164	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111620\
Data File : VL035954.D
Acq On : 16 Nov 2020 21:52
Operator : SY/AP
Sample : L4722-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SR-4DL

Manual Integrations
APPROVED

MMDadoda
11/18/2020 9:07:52 AM

Quant Time: Nov 17 22:14:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

