

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111620\
 Data File : VL035965.D
 Acq On : 17 Nov 2020 5:01
 Operator : SY/AP
 Sample : L4773-02 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2

Manual Integrations
 APPROVED

MMDadoda
 11/18/2020 9:08:38 AM

Quant Time: Nov 17 22:34:40 2020
 Quant Title :
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1200208	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	4465320	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	4474518	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	3414715	10.74	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.01	41	15730	0.299	ppbv	88
13) Ethanol	3.61	45	86594m	7.588	ppbv	
15) Acetone	3.85	43	927788	7.231	ppbv	96
19) Isopropyl Alcohol	3.98	45	35864	0.455	ppbv #	90
26) Hexane	5.69	57	234652	1.683	ppbv #	46
29) Chloroform	5.75	83	203140	0.741	ppbv	98
31) cis-1,2-Dichloroethene	5.54	61	30434m	0.212	ppbv	
38) Trichloroethene	7.84	130	563747	3.477	ppbv	95
52) Tetrachloroethene	11.23	164	9543133m	60.678	ppbv	
60) o-Xylene	13.70	91	54667m	0.122	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.68	83	28919m	0.083	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111620\
Data File : VL035965.D
Acq On : 17 Nov 2020 5:01
Operator : SY/AP
Sample : L4773-02 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2

Quant Time: Nov 17 22:34:40 2020
Quant Title :
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
11/18/2020 9:08:38 AM

