


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL111620\
 Data File : VL035937.D
 Acq On : 16 Nov 2020 9:57
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL110320AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
 Last Update : Wed Nov 04 02:40:06 2020



AutoFind: Scans 3618, 3619, 3620; Background Corrected with Scan 3594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	17.4	235370	PASS
75	95	30	66	49.9	672981	PASS
95	95	100	100	100.0	1349973	PASS
96	95	5	9	7.1	96272	PASS
173	174	0.00	2	0.5	4826	PASS
174	95	50	120	70.5	951680	PASS
175	174	4	9	7.3	69005	PASS
176	174	93	101	96.3	916245	PASS
177	176	5	9	6.7	61533	PASS