

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111820\
 Data File : VL035973.D
 Acq On : 18 Nov 2020 14:58
 Operator : SY/AP
 Sample : L4773-02DL2 100X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV2DL2

Manual Integrations
 APPROVED

MMDadoda
 11/20/2020 12:27:19 PM

Quant Time: Nov 19 12:35:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 0
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	1064369	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.14	114	3867359	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	3748914	10.00	ppbv	-0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.38	95	2785858	10.46	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.60%
Target Compounds						
38) Trichloroethene	7.79	130	37583m	0.268	ppbv	Ovalue
52) Tetrachloroethene	11.19	164	1427630	10.481	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111820\
Data File : VL035973.D
Acq On : 18 Nov 2020 14:58
Operator : SY/AP
Sample : L4773-02DL2 100X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV2DL2

Manual Integrations
APPROVED

MMDadoda
11/20/2020 12:27:19 PM

Quant Time: Nov 19 12:35:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

