

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111820\
 Data File : VL035978.D
 Acq On : 18 Nov 2020 18:08
 Operator : SY/AP
 Sample : L4773-01DL2 400X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL2

Manual Integrations
 APPROVED

MMDadoda
 11/20/2020 12:27:28 PM

Quant Time: Nov 19 08:57:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 04 02:40:06 2020
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	1108784	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.13	114	3921384	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.04	117	3760736	10.00	ppbv	-0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.37	95	2774435	10.38	ppbv	-0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.59	45	11540m	1.095	ppbv	
15) Acetone	3.84	43	43782m	0.369	ppbv	
20) Methylene Chloride	4.31	84	21684	0.260	ppbv	92
29) Chloroform	5.71	83	30335	0.120	ppbv	89
30) Cyclohexane	7.12	84	25785m	0.184	ppbv	
38) Trichloroethene	7.79	130	23777	0.167	ppbv	97
52) Tetrachloroethene	11.18	164	996806	7.217	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111820\
Data File : VL035978.D
Acq On : 18 Nov 2020 18:08
Operator : SY/AP
Sample : L4773-01DL2 400X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL2

Manual Integrations
APPROVED

MMDadoda
11/20/2020 12:27:28 PM

Quant Time: Nov 19 08:57:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

