

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111820\
 Data File : VL035979.D
 Acq On : 18 Nov 2020 18:45
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 11/20/2020 12:27:31 PM

Quant Time: Nov 19 09:00:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 0
 QLast Update : Wed Nov 04 02:40:06 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	1129659	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.13	114	3960995	10.00	ppbv	-0.05
55) Chlorobenzene-d5	12.04	117	3682220	10.00	ppbv	-0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.37	95	2752573	10.52	ppbv	-0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.59	45	6032m	0.562	ppbv	
15) Acetone	3.83	43	40882m	0.339	ppbv	
20) Methylene Chloride	4.31	84	95379	1.122	ppbv	92
26) Hexane	5.65	57	27794m	0.212	ppbv	
30) Cyclohexane	7.12	84	24130m	0.169	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111820\
Data File : VL035979.D
Acq On : 18 Nov 2020 18:45
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
11/20/2020 12:27:31 PM

Quant Time: Nov 19 09:00:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL110320AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 04 02:40:06 2020
Response via : Initial Calibration

