

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111919\
Data File : VL034384.D
Acq On : 19 Nov 2019 18:11
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Nov 20 05:40:23 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 20 03:30:54 2019
QLast Update : Wed Nov 20 03:30:54 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1033159	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	2027186	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.14	117	1958214	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1691694	10.27	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.70%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111919\
Data File : VL034384.D
Acq On : 19 Nov 2019 18:11
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Nov 20 05:40:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 20 03:30:54 2019
QLast Update : Wed Nov 20 03:30:54 2019
Response via : Initial Calibration

