

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL111919\
Data File : VL034388.D
Acq On : 19 Nov 2019 20:54
Operator : SY/AP
Sample : OC110719 CAN 10449
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC110719 CAN 10449

Quant Time: Nov 20 11:05:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 20 03:30:54 2019
QLast Update : Wed Nov 20 03:30:54 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	949251	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	1836990	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.14	117	1827116	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1557833	10.13	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.30%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL111919\
Data File : VL034388.D
Acq On : 19 Nov 2019 20:54
Operator : SY/AP
Sample : OC110719 CAN 10449
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC110719 CAN 10449

Quant Time: Nov 20 11:05:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 20 03:30:54 2019
QLast Update : Wed Nov 20 03:30:54 2019
Response via : Initial Calibration

