

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112018\
Data File : VL032782.D
Acq On : 20 Nov 2018 17:26
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Nov 21 05:41:58 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 14 07:00:23 2018
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1403619	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3514093	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3155994	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2341790	9.92	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.20%

Target Compounds					Ovalue
52) Tetrachloroethene	11.38	164	233490	2.399	ppbv 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112018\
Data File : VL032782.D
Acq On : 20 Nov 2018 17:26
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Nov 21 05:41:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 07:00:23 2018
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

