

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112018\
 Data File : VL032801.D
 Acq On : 21 Nov 2018 5:51
 Operator : SY/AP
 Sample : J6015-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL

Manual Integrations
 APPROVED

MMDadoda
 11/26/2018 9:19:03 AM

Quant Time: Nov 21 07:14:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Nov 14 07:14:03 2018
 QLast Update : Wed Nov 14 07:00:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1331918	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3417017	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3039101	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2379815	10.47	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.66	45	119356	8.623	ppbv	98
15) Acetone	3.92	43	53205	0.437	ppbv	92
19) Isopropyl Alcohol	4.05	45	20480m	0.253	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112018\
Data File : VL032801.D
Acq On : 21 Nov 2018 5:51
Operator : SY/AP
Sample : J6015-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL

Quant Time: Nov 21 07:14:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

Manual Integrations
APPROVED

MMDadoda
11/26/2018 9:19:03 AM

