

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112018\
Data File : VL032777.D
Acq On : 20 Nov 2018 13:35
Operator : SY/AP
Sample : VL1120ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1120ABL01

Quant Time: Nov 21 04:17:54 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Nov 14 07:00:23 2018
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1418857	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3691954	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3304094	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2403073	9.72	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.20%


Target Compounds	Ovalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112018\
Data File : VL032777.D
Acq On : 20 Nov 2018 13:35
Operator : SY/AP
Sample : VL1120ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1120ABL01

Quant Time: Nov 21 04:17:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov 14 07:00:23 2018
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.000 ppbv

RT: 5.77 min Scan# 882

Delta R.T. -0.00 min

Lab File: VL032777.D

Acq: 20 Nov 2018 13:35

Instrument :


MSVOA_L

ClientSampleId :

VL1120ABL01

Tot Ion: 49 Resp: 1418857



Ion	Ratio	Lower	Upper
49	100		
128	53.7	24.9	74.6
130	69.6	31.4	94.2



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): V

Ion 130.00 (129.70 to 130.70): V



#33

1,4-Difluorobenzene

Concen: 10.000 ppbv

RT: 7.30 min Scan# 1357


Delta R.T. -0.00 min

Lab File: VL032777.D

Acq: 20 Nov 2018 13:35

Tgt Ion:114 Resp: 3691954


Ion	Ratio	Lower	Upper
114	100		
63	24.3	20.9	31.3
88	17.7	14.4	21.6




Abundance Ion 114.10 (113.80 to 114.80): V

Ion 63.10 (62.80 to 63.80): VL0

Ion 88.10 (87.80 to 88.80): VL0









#55
Chlorobenzene-d5
Concen: 10.000 ppbv
RT: 12.25 min Scan# 2896
Delta R.T. -0.00 min
Lab File: VL032777.D
Acq: 20 Nov 2018 13:35

Instrument :
MSVOA_L
ClientSampleId :
VL1120ABL01

Tot Ion:117 Resp: 3304094
Ion Ratio Lower Upper
117 100
82 64.3 51.0 76.6
119 32.8 28.0 42.0



#68
1-Bromo-4-Fluorobenzene
Concen: 9.724 ppbv
RT: 14.59 min Scan# 3624
Delta R.T. -0.00 min
Lab File: VL032777.D
Acq: 20 Nov 2018 13:35

Tgt Ion: 95 Resp: 2403073
Ion Ratio Lower Upper
95 100
174 68.0 50.7 76.1
175 5.0 3.8 5.6

