

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112024\
 Data File : VL041708.D
 Acq On : 20 Nov 2024 20:33
 Operator : SY/MD
 Sample : P4911-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
 Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 05:15:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Nov 08 01:01:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.211	49	1198174	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.674	114	3245825	10.000	ppbv	0.02
55) Chlorobenzene-d5	11.487	117	2859554	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.796	95	2092511	9.931	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.300%
Target Compounds						Qvalue
9) Propene	2.723	41	63961	0.940	ppbv	86
13) Ethanol	3.279	45	13896m	3.336	ppbv	
26) Hexane	5.230	57	60214	0.513	ppbv #	29
34) 2-Butanone	4.825	43	52579	0.325	ppbv	95
62) Isopropylbenzene	14.050	105	148209	0.358	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112024\
Data File : VL041708.D
Acq On : 20 Nov 2024 20:33
Operator : SY/MD
Sample : P4911-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/21/2024
Supervised By :Mahesh Dadoda 11/21/2024

Quant Time: Nov 21 05:15:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL110724AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Nov 08 01:01:00 2024
Response via : Initial Calibration

