

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112122\
 Data File : VL039912.D
 Acq On : 21 Nov 2022 16:25
 Operator : SY/AP
 Sample : N5724-05DL 4X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/22/2022
 Supervised By :Mahesh Dadoda 11/24/2022

Quant Time: Nov 22 03:49:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Tue Nov 22 03:45:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.166	49	1142101	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.629	114	3122356	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.449	117	2801590	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.764	95	2230172	10.422	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.200%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.706	85	20144	0.123	ppbv	94
9) Propene	2.684	41	11214m	0.151	ppbv	
13) Ethanol	3.234	45	22653m	3.997	ppbv	
15) Acetone	3.452	43	139396	0.920	ppbv	99
20) Methylene Chloride	3.909	84	791142	10.780	ppbv	98
26) Hexane	5.188	57	421653	3.470	ppbv #	38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL112122\
Data File : VL039912.D
Acq On : 21 Nov 2022 16:25
Operator : SY/AP
Sample : N5724-05DL 4X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA1DL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/22/2022
Supervised By :Mahesh Dadoda 11/24/2022

Quant Time: Nov 22 03:49:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Tue Nov 22 03:45:31 2022
Response via : Initial Calibration

