

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL112122\  
 Data File : VL039905.D  
 Acq On : 21 Nov 2022 10:39  
 Operator : SY/AP  
 Sample : VSTDCCC010  
 Misc : 400mL/MSVOA\_L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 LabSampleId :  
 VSTDCCC010

Quant Time: Nov 22 03:42:48 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_L\methods\VL111822AIR.M  
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Fri Aug 26 06:05:16 2022  
 QLast Update : Fri Nov 18 14:05:14 2022  
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF   | %Dev  | Area% | Dev(min) |
|------|-----------------------------|-------|--------|-------|-------|----------|
| 1 I  | Bromochloromethane          | 1.000 | 1.000  | 0.0   | 87    | 0.00     |
| 2 T  | Dichlorodifluoromethane     | 1.432 | 1.101  | 23.1  | 90    | 0.00     |
| 3    | Chlorodifluoromethane       | 1.380 | 1.320  | 4.3   | 88    | 0.00     |
| 4    | Chloromethane               | 0.882 | 0.851  | 3.5   | 85    | 0.00     |
| 5 T  | Vinyl Chloride              | 0.689 | 0.737  | -7.0  | 85    | 0.00     |
| 6 T  | Bromomethane                | 0.235 | 0.237  | -0.9  | 86    | 0.00     |
| 7    | Chloroethane                | 0.272 | 0.272  | 0.0   | 88    | 0.00     |
| 8 T  | Dichlorotetrafluoroethane   | 1.635 | 1.601  | 2.1   | 91    | 0.00     |
| 9 T  | Propene                     | 0.650 | 0.623  | 4.2   | 84    | 0.00     |
| 10 T | Heptane                     | 1.568 | 1.537  | 2.0   | 87    | 0.00     |
| 11 T | Trichlorofluoromethane      | 1.747 | 1.836  | -5.1  | 95    | 0.00     |
| 12 T | 1,1,2-Trichlorotrifluoroeth | 1.322 | 1.382  | -4.5  | 94    | 0.00     |
| 13   | Ethanol                     | 0.050 | 0.044# | 12.0  | 88    | 0.00     |
| 14 T | Bromoethene                 | 0.516 | 0.551  | -6.8  | 90    | 0.00     |
| 15 T | Acetone                     | 1.326 | 1.276  | 3.8   | 88    | 0.00     |
| 16 T | 1,3-Butadiene               | 0.661 | 0.679  | -2.7  | 84    | 0.00     |
| 17   | tert-Butyl alcohol          | 1.525 | 1.568  | -2.8  | 87    | 0.00     |
| 18 T | 1,1-Dichloroethene          | 0.603 | 0.626  | -3.8  | 92    | 0.00     |
| 19 T | Isopropyl Alcohol           | 0.718 | 0.734  | -2.2  | 85    | 0.00     |
| 20 T | Methylene Chloride          | 0.643 | 0.526  | 18.2  | 91    | 0.00     |
| 21 T | Allyl Chloride              | 0.920 | 0.952  | -3.5  | 87    | 0.00     |
| 22 T | trans-1,2-Dichloroethene    | 0.613 | 0.655  | -6.9  | 93    | 0.00     |
| 23 T | Vinyl Acetate               | 0.545 | 0.548  | -0.6  | 86    | 0.00     |
| 24 T | 1,1-Dichloroethane          | 1.314 | 1.348  | -2.6  | 91    | 0.00     |
| 25 T | Ethyl Acetate               | 2.435 | 2.456  | -0.9  | 88    | 0.00     |
| 26 T | Hexane                      | 1.064 | 1.035  | 2.7   | 87    | 0.00     |
| 27 T | Carbon Disulfide            | 1.343 | 1.568  | -16.8 | 89    | 0.00     |
| 28 T | Methyl tert-Butyl Ether     | 1.332 | 1.358  | -2.0  | 90    | 0.00     |
| 29 T | Chloroform                  | 1.414 | 1.561  | -10.4 | 91    | 0.00     |
| 30 T | Cyclohexane                 | 1.044 | 1.034  | 1.0   | 89    | 0.00     |
| 31 T | cis-1,2-Dichloroethene      | 0.983 | 1.044  | -6.2  | 89    | 0.00     |
| 32 T | 1,1,1-Trichloroethane       | 1.766 | 1.953  | -10.6 | 97    | 0.00     |
| 33 I | 1,4-Difluorobenzene         | 1.000 | 1.000  | 0.0   | 90    | 0.00     |
| 34 T | 2-Butanone                  | 0.478 | 0.504  | -5.4  | 88    | 0.00     |
| 35 T | Carbon Tetrachloride        | 0.568 | 0.666  | -17.3 | 96    | 0.00     |
| 36 T | Benzene                     | 0.850 | 0.862  | -1.4  | 90    | 0.00     |
| 37 T | 1,2-Dichloroethane          | 0.466 | 0.494  | -6.0  | 94    | 0.00     |
| 38 T | Trichloroethene             | 0.475 | 0.487  | -2.5  | 93    | 0.00     |
| 39 T | 1,2-Dichloropropane         | 0.331 | 0.329  | 0.6   | 87    | 0.00     |
| 40 T | 1,4-Dioxane                 | 0.132 | 0.134  | -1.5  | 91    | 0.00     |
| 41 T | Tetrahydrofuran             | 0.330 | 0.329  | 0.3   | 83    | 0.00     |
| 42 T | Bromodichloromethane        | 0.632 | 0.715  | -13.1 | 93    | 0.00     |
| 43   | Methyl Methacrylate         | 0.333 | 0.349  | -4.8  | 87    | 0.00     |
| 44 T | 2,2,4-Trimethylpentane      | 1.419 | 1.380  | 2.7   | 87    | 0.00     |
| 45 T | t-1,3-Dichloropropene       | 0.308 | 0.368  | -19.5 | 88    | 0.00     |
| 46 T | cis-1,3-Dichloropropene     | 0.427 | 0.470  | -10.1 | 89    | 0.00     |
| 47 T | 1,1,2-Trichloroethane       | 0.340 | 0.352  | -3.5  | 92    | 0.00     |
| 48 T | Dibromochloromethane        | 0.499 | 0.605  | -21.2 | 94    | 0.00     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_L\Data\VL112122\  
 Data File : VL039905.D  
 Acq On : 21 Nov 2022 10:39  
 Operator : SY/AP  
 Sample : VSTDCCC010  
 Misc : 400mL/MSVOA\_L  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 MSVOA\_L  
 LabSampleId :  
 VSTDCCC010

Quant Time: Nov 22 03:42:48 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_L\methods\VL111822AIR.M  
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA\_L Fri Aug 26 06:05:16 2022  
 QLast Update : Fri Nov 18 14:05:14 2022  
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 30% Max. Rel. Area : 150%

|      | Compound                  | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|---------------------------|-------|-------|-------|-------|----------|
| 49 T | Bromoform                 | 0.404 | 0.491 | -21.5 | 88    | 0.00     |
| 50 T | 4-Methyl-2-Pentanone      | 0.857 | 0.868 | -1.3  | 86    | 0.00     |
| 51 T | 2-Hexanone                | 0.665 | 0.699 | -5.1  | 87    | 0.00     |
| 52 T | Tetrachloroethene         | 0.318 | 0.308 | 3.1   | 88    | 0.00     |
| 53 T | Toluene                   | 0.968 | 0.982 | -1.4  | 90    | 0.00     |
| 54 T | 1,2-Dibromoethane         | 0.363 | 0.414 | -14.0 | 94    | 0.00     |
| 55 I | Chlorobenzene-d5          | 1.000 | 1.000 | 0.0   | 88    | 0.00     |
| 56   | 1,1,1,2-Tetrachloroethane | 0.409 | 0.442 | -8.1  | 93    | 0.00     |
| 57 T | Chlorobenzene             | 0.765 | 0.786 | -2.7  | 92    | 0.00     |
| 58 T | Ethyl Benzene             | 1.357 | 1.361 | -0.3  | 90    | 0.00     |
| 59 T | m/p-Xylene                | 1.135 | 1.142 | -0.6  | 92    | 0.00     |
| 60 T | o-Xylene                  | 1.083 | 1.093 | -0.9  | 91    | 0.00     |
| 61 T | Styrene                   | 0.646 | 0.688 | -6.5  | 90    | 0.00     |
| 62   | Isopropylbenzene          | 1.562 | 1.558 | 0.3   | 92    | 0.00     |
| 63 T | 1,1,2,2-Tetrachloroethane | 0.362 | 0.373 | -3.0  | 91    | 0.00     |
| 64   | n-propylbenzene           | 0.408 | 0.422 | -3.4  | 91    | 0.00     |
| 65   | tert-Butylbenzene         | 1.366 | 1.357 | 0.7   | 92    | 0.00     |
| 66 T | Benzyl Chloride           | 0.117 | 0.144 | -23.1 | 83    | 0.00     |
| 67   | sec-Butylbenzene          | 1.878 | 1.876 | 0.1   | 92    | 0.00     |
| 68 S | 1-Bromo-4-Fluorobenzene   | 0.764 | 0.758 | 0.8   | 89    | 0.00     |
| 69   | p-Isopropyltoluene        | 1.547 | 1.543 | 0.3   | 92    | 0.00     |
| 70   | n-Butylbenzene            | 1.534 | 1.550 | -1.0  | 92    | 0.00     |
| 71   | 2-Chlorotoluene           | 1.184 | 1.189 | -0.4  | 91    | 0.00     |
| 72 T | 4-Ethyltoluene            | 1.217 | 1.279 | -5.1  | 92    | 0.00     |
| 73 T | 1,3,5-Trimethylbenzene    | 1.114 | 1.135 | -1.9  | 92    | 0.00     |
| 74 T | 1,2,4-Trimethylbenzene    | 1.180 | 1.196 | -1.4  | 92    | 0.00     |
| 75 T | 1,3-Dichlorobenzene       | 0.664 | 0.689 | -3.8  | 90    | 0.00     |
| 76 T | 1,4-Dichlorobenzene       | 0.658 | 0.668 | -1.5  | 88    | 0.00     |
| 77 T | 1,2-Dichlorobenzene       | 0.640 | 0.649 | -1.4  | 90    | 0.00     |
| 78 T | Hexachloro-1,3-Butadiene  | 0.444 | 0.392 | 11.7  | 86    | 0.00     |
| 79 T | Naphthalene               | 0.706 | 0.785 | -11.2 | 90    | 0.00     |
| 80 T | Naphthalene,2-methyl-     | 0.199 | 0.243 | -22.1 | 106   | 0.00     |
| 81 T | 1,2,4-Trichlorobenzene    | 0.436 | 0.436 | 0.0   | 88    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0