

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112321\
 Data File : VL038114.D
 Acq On : 24 Nov 2021 00:27
 Operator : SY/AP
 Sample : M4805-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA L
 ClientSampleId :
 SG-ROOK-COFFEEDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/29/2021
 Supervised By :Mahesh Dadoda 11/29/2021

Quant Time: Nov 28 12:05:55 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	824306	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	2398148	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	2096103	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	1632211	10.80	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	108.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.62	45	37882m	6.149	ppbv	
15) Acetone	3.87	43	150387m	2.104	ppbv	
19) Isopropyl Alcohol	4.00	45	20786m	0.403	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112321\
Data File : VL038114.D
Acq On : 24 Nov 2021 00:27
Operator : SY/AP
Sample : M4805-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SG-ROOK-COFFEEDL

Manual Integrations
APPROVED

Quant Time: Nov 28 12:05:55 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 11/29/2021
Supervised By :Mahesh Dadoda 11/29/2021

