

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112519\
 Data File : VL034416.D
 Acq On : 25 Nov 2019 21:37
 Operator : SY/AP
 Sample : K5971-05DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 5DL

Manual Integrations
 APPROVED

apatel
 12/2/2019 10:04:30 AM

Quant Time: Nov 26 04:22:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Nov 19 14:15:35 2019
 QLast Update : Tue Nov 19 14:15:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	864171	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	1703146	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	1667739	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1429857	10.19	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	208727	15.653	ppbv	98
15) Acetone	3.84	43	114533	0.926	ppbv	98
20) Methylene Chloride	4.34	84	17094m	0.390	ppbv	
26) Hexane	5.69	57	16533	0.174	ppbv #	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112519\
Data File : VL034416.D
Acq On : 25 Nov 2019 21:37
Operator : SY/AP
Sample : K5971-05DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
5DL

Manual Integrations
APPROVED

apatel
12/2/2019 10:04:30 AM

Quant Time: Nov 26 04:22:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov
QLast Update : Tue Nov 19 14:15:35 2019
Response via : Initial Calibration

