

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL112718\
 Data File : VL032809.D
 Acq On : 27 Nov 2018 16:13
 Operator : SY/AP
 Sample : VL1127ABL01
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL1127ABL01

Manual Integrations
 APPROVED

MMDadoda
 11/28/2018 4:21:23 PM

Quant Time: Nov 28 03:30:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL111418AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Nov 14 07:00:23 2018
 QLast Update : Wed Nov 14 07:00:23 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.77	49	1302953	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.30	114	3390187	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	3177143	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.59	95	2294173	9.65	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
15) Acetone	3.94	43	24478m	0.21	ppbv	
20) Methylene Chloride	4.42	84	17147m	0.28	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL112718\
Data File : VL032809.D
Acq On : 27 Nov 2018 16:13
Operator : SY/AP
Sample : VL1127ABL01
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL1127ABL01

Manual Integrations
APPROVED

MMDadoda
11/28/2018 4:21:23 PM

Quant Time: Nov 28 03:30:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL111418AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Nov
QLast Update : Wed Nov 14 07:00:23 2018
Response via : Initial Calibration

