

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL112921\
 Data File : VL038134.D
 Acq On : 29 Nov 2021 14:18
 Operator : SY/AP
 Sample : M4751-02DL2 1200X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SVP-2DL2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 11/30/2021
 Supervised By :Mahesh Dadoda 12/01/2021

Quant Time: Nov 30 21:19:57 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	948488	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2539691	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.09	117	2222848	10.00	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.42	95	1757408	10.96	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	109.60%
Target Compounds						
38) Trichloroethene	7.84	130	25260m	0.264	ppbv	Ovalue
52) Tetrachloroethene	11.23	164	1045851	11.157	ppbv	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL112921\
Data File : VL038134.D
Acq On : 29 Nov 2021 14:18
Operator : SY/AP
Sample : M4751-02DL2 1200X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SVP-2DL2

Quant Time: Nov 30 21:19:57 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 11/30/2021
Supervised By :Mahesh Dadoda 12/01/2021

