

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
 Data File : VL039944.D
 Acq On : 1 Dec 2022 16:23
 Operator : SY/MD
 Sample : N5824-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 TPC-1ADL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/02/2022
 Supervised By :Mahesh Dadoda 12/05/2022

Quant Time: Dec 02 05:33:16 2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 12 2022

QLast Update : Fri Dec 02 05:29:35 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.166	49	1292659	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.626	114	3530131	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.442	117	3187443	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.754	95	2561430	10.521	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	105.200%
Target Compounds						
					Qvalue	
9) Propene	2.684	41	12026m	0.143	ppbv	
13) Ethanol	3.208	45	1229569m	191.669	ppbv	
15) Acetone	3.446	43	304809	1.778	ppbv	96
19) Isopropyl Alcohol	3.552	45	765311	8.244	ppbv #	32
26) Hexane	5.188	57	107509	0.782	ppbv #	53
53) Toluene	9.198	91	53344	0.156	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
Data File : VL039944.D
Acq On : 1 Dec 2022 16:23
Operator : SY/MD
Sample : N5824-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
TPC-1ADL

Quant Time: Dec 02 05:33:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Dec 02 05:29:35 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/02/2022
Supervised By : Mahesh Dadoda 12/05/2022

