

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
 Data File : VL039946.D
 Acq On : 1 Dec 2022 17:42
 Operator : SY/MD
 Sample : N5825-01DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OTC-IADL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/02/2022
 Supervised By :Mahesh Dadoda 12/05/2022

Quant Time: Dec 02 05:33:50 2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 12 2022

QLast Update : Fri Dec 02 05:29:35 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.163	49	1274322	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	6.623	114	3436666	10.000	ppbv	-0.01
55) Chlorobenzene-d5	11.436	117	3107868	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.751	95	2540672	10.703	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	107.000%
Target Compounds						
3) Chlorodifluoromethane	2.645	51	203091	1.155	ppbv	98
9) Propene	2.671	41	13008m	0.157	ppbv	
13) Ethanol	3.208	45	635679m	100.518	ppbv	
15) Acetone	3.443	43	137623	0.814	ppbv	96
19) Isopropyl Alcohol	3.555	45	82778m	0.905	ppbv	
20) Methylene Chloride	3.906	84	48594	0.593	ppbv #	80
25) Ethyl Acetate	5.179	43	88575	0.285	ppbv	98
26) Hexane	5.182	57	45860	0.338	ppbv	87
52) Tetrachloroethene	10.581	164	3810m	0.035	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
Data File : VL039946.D
Acq On : 1 Dec 2022 17:42
Operator : SY/MD
Sample : N5825-01DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OTC-IADL

Quant Time: Dec 02 05:33:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL11822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Dec 02 05:29:35 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/02/2022
Supervised By : Mahesh Dadoda 12/05/2022

