

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
 Data File : VL039952.D
 Acq On : 1 Dec 2022 21:36
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/02/2022
 Supervised By :Mahesh Dadoda 12/05/2022

Quant Time: Dec 02 05:35:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Fri Dec 02 05:29:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.160	49	1155356	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	6.619	114	3217147	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.436	117	2910503	10.000	ppbv	-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.748	95	2292062	10.310	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.100%
Target Compounds						
52) Tetrachloroethene	10.590	164	6818m	0.067	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
Data File : VL039952.D
Acq On : 1 Dec 2022 21:36
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/02/2022
Supervised By : Mahesh Dadoda 12/05/2022

Quant Time: Dec 02 05:35:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Dec 02 05:29:35 2022
Response via : Initial Calibration

