

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
 Data File : VL039958.D
 Acq On : 2 Dec 2022 1:26
 Operator : SY/MD
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Quant Time: Dec 02 05:37:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 05:29:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.160	49	1055605	10.000	ppbv	-0.01
33) 1,4-Difluorobenzene	6.619	114	3025995	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.436	117	2694505	10.000	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.744	95	2155017	10.471	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
Data File : VL039958.D
Acq On : 2 Dec 2022 1:26
Operator : SY/MD
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Dec 02 05:37:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Fri Dec 02 05:29:35 2022
Response via : Initial Calibration

