

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
 Data File : VL039961.D
 Acq On : 2 Dec 2022 8:28
 Operator : SY/MD
 Sample : N5825-02DL2 200X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OTC-SSDL2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 12/02/2022
 Supervised By : Mahesh Dadoda 12/05/2022

Quant Time: Dec 02 12:07:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug :
 QLast Update : Fri Dec 02 05:29:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.150	49	1020749	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.613	114	3153598	10.000	ppbv	-0.02
55) Chlorobenzene-d5	11.426	117	2853919	10.000	ppbv	-0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.738	95	2216849	10.170	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.700%
Target Compounds						
					Qvalue	
13) Ethanol	3.214	45	13606m	2.686	ppbv	
15) Acetone	3.443	43	240268	1.775	ppbv	99
20) Methylene Chloride	3.896	84	40716	0.621	ppbv	91
26) Hexane	5.176	57	28193m	0.260	ppbv	
52) Tetrachloroethene	10.574	164	360280	3.593	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120122\
Data File : VL039961.D
Acq On : 2 Dec 2022 8:28
Operator : SY/MD
Sample : N5825-02DL2 200X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OTC-SSDL2

Quant Time: Dec 02 12:07:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Dec 02 05:29:35 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 12/02/2022
Supervised By : Mahesh Dadoda 12/05/2022

