

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120219\
 Data File : VL034435.D
 Acq On : 2 Dec 2019 20:45
 Operator : SY/AP
 Sample : K6068-04DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-4DL

Manual Integrations
 APPROVED

apatel
 12/4/2019 4:47:17 PM

Quant Time: Dec 03 11:30:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Nov 19 14:15:35 2019
 QLast Update : Tue Nov 19 14:15:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.69	49	864272	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.21	114	1750717	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.15	117	1671280	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.48	95	1411867	10.04	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	47543m	0.993	ppbv	
13) Ethanol	3.60	45	26108	1.958	ppbv	98
15) Acetone	3.84	43	911134	7.362	ppbv	99
29) Chloroform	5.77	83	24766	0.161	ppbv	91
31) cis-1,2-Dichloroethene	5.56	61	147118	1.500	ppbv	94
32) 1,1,1-Trichloroethane	6.54	97	62411	0.433	ppbv	99
38) Trichloroethene	7.88	130	5206336	72.492	ppbv	95
52) Tetrachloroethene	11.28	164	2710519	35.796	ppbv	96
53) Toluene	9.86	91	46542	0.234	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120219\
Data File : VL034435.D
Acq On : 2 Dec 2019 20:45
Operator : SY/AP
Sample : K6068-04DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-4DL

Manual Integrations
APPROVED

apatel
12/4/2019 4:47:17 PM

Quant Time: Dec 03 11:30:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111919AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LTue Nov
QLast Update : Tue Nov 19 14:15:35 2019
Response via : Initial Calibration

