

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120222\
 Data File : VL039968.D
 Acq On : 2 Dec 2022 13:36
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Dec 02 23:23:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	1.432	1.529	-6.8	112	0.00
3	Chlorodifluoromethane	1.380	1.110	19.6	66	0.00
4	Chloromethane	0.882	0.616	30.2#	55	0.00
5 T	Vinyl Chloride	0.689	0.583	15.4	61	0.00
6 T	Bromomethane	0.235	0.252	-7.2	82	0.00
7	Chloroethane	0.272	0.209	23.2	61	0.00
8 T	Dichlorotetrafluoroethane	1.635	1.473	9.9	75	0.00
9 T	Propene	0.650	0.487	25.1	59	0.00
10 T	Heptane	1.568	1.220	22.2	62	0.00
11 T	Trichlorofluoromethane	1.747	1.542	11.7	72	0.00
12 T	1,1,2-Trichlorotrifluoroeth	1.322	1.153	12.8	70	0.00
13	Ethanol	0.050	0.031#	38.0#	54	0.00
14 T	Bromoethene	0.516	0.461	10.7	68	0.00
15 T	Acetone	1.326	0.987	25.6	61	0.00
16 T	1,3-Butadiene	0.661	0.534	19.2	59	0.00
17	tert-Butyl alcohol	1.525	1.175	23.0	58	0.00
18 T	1,1-Dichloroethene	0.603	0.503	16.6	66	0.00
19 T	Isopropyl Alcohol	0.718	0.506	29.5	52	0.00
20 T	Methylene Chloride	0.643	0.421	34.5#	65	0.00
21 T	Allyl Chloride	0.920	0.713	22.5	59	0.00
22 T	trans-1,2-Dichloroethene	0.613	0.525	14.4	67	0.00
23 T	Vinyl Acetate	0.545	0.181	66.8#	26	0.00
24 T	1,1-Dichloroethane	1.314	1.068	18.7	65	0.00
25 T	Ethyl Acetate	2.435	2.455	-0.8	79	0.00
26 T	Hexane	1.064	0.917	13.8	69	0.00
27 T	Carbon Disulfide	1.343	1.234	8.1	63	0.00
28 T	Methyl tert-Butyl Ether	1.332	1.030	22.7	61	0.00
29 T	Chloroform	1.414	1.537	-8.7	80	0.00
30 T	Cyclohexane	1.044	0.832	20.3	64	0.00
31 T	cis-1,2-Dichloroethene	0.983	0.825	16.1	63	0.00
32 T	1,1,1-Trichloroethane	1.766	1.598	9.5	71	0.00
33 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
34 T	2-Butanone	0.478	0.412	13.8	61	0.00
35 T	Carbon Tetrachloride	0.568	0.594	-4.6	72	0.00
36 T	Benzene	0.850	0.741	12.8	65	0.00
37 T	1,2-Dichloroethane	0.466	0.422	9.4	68	0.00
38 T	Trichloroethene	0.475	0.359	24.4	58	0.00
39 T	1,2-Dichloropropane	0.331	0.278	16.0	62	0.00
40 T	1,4-Dioxane	0.132	0.119	9.8	68	0.00
41 T	Tetrahydrofuran	0.330	0.284	13.9	60	0.00
42 T	Bromodichloromethane	0.632	0.625	1.1	69	0.00
43	Methyl Methacrylate	0.333	0.313	6.0	66	0.00
44 T	2,2,4-Trimethylpentane	1.419	1.197	15.6	64	0.00
45 T	t-1,3-Dichloropropene	0.308	0.271	12.0	55	0.00
46 T	cis-1,3-Dichloropropene	0.427	0.371	13.1	59	0.00
47 T	1,1,2-Trichloroethane	0.340	0.314	7.6	69	0.00
48 T	Dibromochloromethane	0.499	0.527	-5.6	69	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL120222\
 Data File : VL039968.D
 Acq On : 2 Dec 2022 13:36
 Operator : SY/MD
 Sample : VSTDCCC010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 LabSampleId :
 VSTDCCC010

Quant Time: Dec 02 23:23:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL111822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Fri Dec 02 23:21:36 2022
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	Bromoform	0.404	0.424	-5.0	64	0.00
50 T	4-Methyl-2-Pentanone	0.857	0.743	13.3	62	0.00
51 T	2-Hexanone	0.665	0.583	12.3	61	0.00
52 T	Tetrachloroethene	0.318	0.256	19.5	62	0.00
53 T	Toluene	0.968	0.851	12.1	66	0.00
54 T	1,2-Dibromoethane	0.363	0.404	-11.3	78	0.00
55 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
56	1,1,1,2-Tetrachloroethane	0.409	0.407	0.5	74	0.00
57 T	Chlorobenzene	0.765	0.664	13.2	68	0.00
58 T	Ethyl Benzene	1.357	1.178	13.2	68	0.00
59 T	m/p-Xylene	1.135	0.992	12.6	69	0.00
60 T	o-Xylene	1.083	0.952	12.1	69	0.00
61 T	Styrene	0.646	0.563	12.8	64	0.00
62	Isopropylbenzene	1.562	1.434	8.2	73	0.00
63 T	1,1,2,2-Tetrachloroethane	0.362	0.525	-45.0#	111	0.00
64	n-propylbenzene	0.408	0.376	7.8	70	0.00
65	tert-Butylbenzene	1.366	1.264	7.5	74	0.00
66 T	Benzyl Chloride	0.117	0.069	41.0#	35	0.00
67	sec-Butylbenzene	1.878	1.760	6.3	75	0.00
68 S	1-Bromo-4-Fluorobenzene	0.764	0.768	-0.5	78	0.00
69	p-Isopropyltoluene	1.547	1.447	6.5	74	0.00
70	n-Butylbenzene	1.534	1.485	3.2	76	0.00
71	2-Chlorotoluene	1.184	1.093	7.7	73	0.00
72 T	4-Ethyltoluene	1.217	1.099	9.7	68	0.00
73 T	1,3,5-Trimethylbenzene	1.114	0.984	11.7	69	0.00
74 T	1,2,4-Trimethylbenzene	1.180	1.055	10.6	70	0.00
75 T	1,3-Dichlorobenzene	0.664	0.596	10.2	68	0.00
76 T	1,4-Dichlorobenzene	0.658	0.574	12.8	66	0.00
77 T	1,2-Dichlorobenzene	0.640	0.569	11.1	68	0.00
78 T	Hexachloro-1,3-Butadiene	0.444	0.400	9.9	76	0.00
79 T	Naphthalene	0.706	0.820	-16.1	82	0.00
80 T	Naphthalene,2-methyl-	0.199	0.223	-12.1	84	0.00
81 T	1,2,4-Trichlorobenzene	0.436	0.420	3.7	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0