

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL120721\
 Data File : VL038148.D
 Acq On : 7 Dec 2021 13:31
 Operator : SY/AP
 Sample : M4827-03DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 AA-1DL

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/07/2021
 Supervised By :Semsettin Yesilyurt 12/08/2021

Quant Time: Dec 07 15:09:57 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	908429	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	2388100	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	2056573	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	1642952	11.08	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	110.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.63	45	17002m	2.504	ppbv	
15) Acetone	3.86	43	50969m	0.647	ppbv	
20) Methylene Chloride	4.34	84	146624	3.324	ppbv	87
26) Hexane	5.68	57	77809	0.617	ppbv #	44

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL120721\
Data File : VL038148.D
Acq On : 7 Dec 2021 13:31
Operator : SY/AP
Sample : M4827-03DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
AA-1DL

Manual Integrations
APPROVED

Quant Time: Dec 07 15:09:57 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

Reviewed By :Amit Patel 12/07/2021
Supervised By :Semsettin Yesilyurt 12/08/2021

