

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL120721\
 Data File : VL038150.D
 Acq On : 7 Dec 2021 14:49
 Operator : SY/AP
 Sample : M4827-02DL2 100X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-2DL2

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/08/2021
 Supervised By :Semsettin Yesilyurt 12/08/2021

Quant Time: Dec 08 00:54:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	852919	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	2283635	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.07	117	1947226	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	1579211	11.24	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	112.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.64	45	5931m	0.930	ppbv	
20) Methylene Chloride	4.34	84	39748	0.960	ppbv	92
36) Benzene	6.88	78	21795m	0.096	ppbv	
52) Tetrachloroethene	11.20	164	703228	8.343	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL120721\
Data File : VL038150.D
Acq On : 7 Dec 2021 14:49
Operator : SY/AP
Sample : M4827-02DL2 100X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-2DL2

Quant Time: Dec 08 00:54:47 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 12/08/2021
Supervised By :Semsettin Yesilyurt 12/08/2021

