

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL120721\
 Data File : VL038158.D
 Acq On : 7 Dec 2021 20:02
 Operator : SY/AP
 Sample : M4897-01DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-1DL

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 12/08/2021
 Supervised By :Semsettin Yesilyurt 12/08/2021

Quant Time: Dec 08 02:36:13 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL111821A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Thu Nov 18 13:16:09 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	764452	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	7.16	114	2217683	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.07	117	1893394	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	1588604	11.63	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	116.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.61	45	54904m	9.609	ppbv	
15) Acetone	3.86	43	168299	2.539	ppbv #	86
19) Isopropyl Alcohol	3.98	45	54586m	1.140	ppbv	
20) Methylene Chloride	4.34	84	28942m	0.780	ppbv	
23) Vinyl Acetate	5.07	43	121789m	1.269	ppbv	
25) Ethyl Acetate	5.67	43	116973	0.531	ppbv #	80
26) Hexane	5.68	57	159878m	1.506	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL120721\
Data File : VL038158.D
Acq On : 7 Dec 2021 20:02
Operator : SY/AP
Sample : M4897-01DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-1DL

Quant Time: Dec 08 02:36:13 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL111821.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Thu Nov 18 13:16:09 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 12/08/2021
Supervised By :Semsettin Yesilyurt 12/08/2021

